



FAA Compliance for Somnetics Devices

To Whom It May Concern:

The US Department of Transportation (DOT) Final Rule, “*Nondiscrimination of the Basis of Disability in Air Travel*” (73 FR 27614 which updates Title 14 CFR Part 382), effective May 13, 2009, provides important new requirements for the accommodation of passengers with respiratory assistive devices (Ventilators, Respirators, and CPAP machines).

In line with these requirements, respiratory assistive devices may be used onboard an aircraft, without further testing by the carrier, provided they have been tested for Electromagnetic Compatibility (EMC) in accordance with the current version of RTCA/DO-160, Section 21, Category M.

Somnetics has now successfully completed testing for the following respiratory assistive devices. **The devices listed below comply with RTCA/DO-160F, Section 21, Category M, and are considered FAA compliant.**

| Product Code | Product Description |
|--------------|--------------------------------------|
| 503001 | Transcend Sleep Apnea Therapy System |